

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	22-09-2025 12:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	22-09-2025 12:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	90 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Labour And Employment
विभाग का नाम/Department Name	Na
संगठन का नाम/Organisation Name	Employees State Insurance Corporation (esic)
कार्यालय का नाम/Office Name	Esic Model Hospital Namkum Ranchi Jharkhand
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	Laparoscopic Surgery System (V2) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	45 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	380 Lakh (s)
वर्षों के अनुभव के लिए एमएसई को छूट प्राप्त है / MSE Exemption for Years Of Experience	Yes Complete
वर्षों के अनुभव के लिए स्टार्टअप को छूट प्राप्त है / Startup Exemption for Years Of Experience	Yes Complete
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 4 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)

बिड विवरण/Bid Details	
बिड लगाने की समय-सीमा बढ़ाने के लिए आवश्यक न्यूनतम सहभागी विक्रेताओं की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
विगत प्रदर्शन / Past Performance	50 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required	Yes
बिड का प्रकार/ Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय / Time allowed for Technical Clarifications during technical evaluation	3 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/ Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य / Estimated Bid Value	9500000
मूल्यांकन पद्धति/ Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/ Arbitration Clause	No
सुलह खंड/ Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	250000

ईपीबीजी विवरण /ePBG Detail

आवश्यकता/Required	No
-------------------	----

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be

in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

MEDICAL SUPERINTENDENT

Esic Model Hospital Namkum Ranchi Jharkhand, NA, Employees State Insurance Corporation (ESIC), Ministry of Labour and Employment
(Esic Fund Account No-02)

विभाजन/Splitting

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	Yes
--	-----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking exemption from Experience Criteria, shall upload the supporting documents to prove his eligibility for exemption.
2. If the bidder is a DPIIT registered Startup, the bidder shall be exempted from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking exemption from Experience Criteria, shall upload the supporting documents to prove his eligibility for exemption.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
5. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

7. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Laparoscopic Surgery System (V2) (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	Laparoscopic Surgery System
	Components of the laparoscopic set to be from the same OEM	Camera head, Camera console, Light source and cable, Telescope
CAMERA PROCESSOR	Camera processor	2D, 3D
	Resolution of camera in pixels	3840 x 2160 (4K)
	Aspect ratio	16:9
	Image sensor	CMOS/CCD
	Progressive scanning system for camera	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Special filters for chrome endoscopy	Yes Or higher
	Minimum Video output	2
	Minimum Number of programmable buttons on camera head	3
	Zoom facility	Yes
	The system capable of ICG fluoroscopy for better delineation of anatomic structures	Yes Or higher
	The camera head shall be sterilizable by autoclave or by some other appropriate technology	Yes
	Camera shall have excellent performance under low light	Yes
	Provision to add fiberoptic filter	Yes
	Should be compatible with flexible video endoscope	Yes, No
	Power supply	100-240V, 50/60 Hz Single phase
HIGH-RESOLUTION MEDICAL GRADE MONITOR	Monitor type	2D, 3D
	Size of monitor in inches	31 to 40, More than 40
	Resolution of monitor in pixels	1920 x 1080 (Full HD), 3840 x 2160 (4K)
	Aspect ratio of monitor	16:9
	Panel shall be anti glare	Yes
	Monitor shall have picture in picture facility	Yes
	Provision of cooling	Yes
	Provision of monitor stand with integrated cable channel for monitor	Yes
	Minimum Video inputs	2
	Minimum Video output of monitor	2
	Viewing angle of monitor	178 degree (H) and 178 degree (V)
	Monitor should be PAL system compatible	Yes
	Availability of output ports	Digital
	Power Supply	100 - 240 VAC 50/60 Hz

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
LIGHT SOURCE	Light Source Type	LED Lamp
	Lamp life	20000 hours or more Or higher
	Colour temperature of light source(Approx)	5401 K to 6500 K Or higher
	Minimum LED Lamp intensity	1800 Lumens or more Or higher
	Means of light intensity adjustment facility	Continuously adjustable from 0 to 100% manually/automatically
	NIR filter for ICG application	Yes
FIBRE OPTIC LIGHT CABLE	Diameter in mm	4.5 mm or above
	Length in cm	300 cm or above
	Cable shall be autoclavable	Yes
	Fibre cable shall be capable of transmitting the optimum light delivered by the light source	Yes
	Cable shall have capability to with stand the high temperature	Yes
CO2 INSUFFLATOR	CO2 insufflator provided	Yes
	Flow rate in Litre per minute	40 or more
	Pressure in mm Hg	0 to 30 mmHg
	Pre heating facility	Yes
	Digital Display for actual and pre set flow rate and pressure and total volume consumed	Yes
	Machine should be able to set flow rate automatically depending on actual abdomen pressure and set pressure	Yes
	Provision of automatic venting mechanism	Yes
	High pressure tubing with yoke assembly from machine to CO2 cylinder	Yes
	Provision of patient tube with luer lock adaptor	Yes
	Availability of hydrophobic filter	Yes
	CO2 insufflator should have visual and audible alarms	Yes
	It should also have provision of automatic feed back on malfunction & indicators on control panel	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
TELESOPES	Compatibility of telescope	HD, 4K
DOCUMENTATION SYSTEM (RECORDER SYSTEM)	Minimum internal memory	1 TB, 2TB Or higher
CERTIFICATIONS	Certification for manufacturing unit	ISO:13485 (Latest)
	Conformity to standards for safety	IEC/EN 60601-1 or equivalent BIS Standard
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	2, 3, 5 Or higher (year)

व्यापक रखरखाव / Comprehensive Maintenance	
Warranty of required product	5 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

परिषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परिषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr Sunita Minz	834010,ESIC MODEL HOSPITAL, NAMKUM, RANCHI	1	30

Special terms and conditions-Version:1 effective from 31-05-2024 for category Laparoscopic Surgery System (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.

3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

3. **Inspection**

Pre-dispatch inspection at Seller premises (Fee/Charges to be borne by the BUYER): Before dispatch, the goods will be inspected by Buyer / Consignee or their Authorized Representative or by Nominated External Inspection Agency (independently or jointly with Buyer or Consignee as decided by the Buyer) at Seller premises (or at designated place for inspection as declared / communicated by the seller) for their compliance to the contract specifications. Fee/Charges taken by the External inspection Agency and any external laboratories testing charges shall be borne by the Buyer. For in-house testing, the Sellers will provide necessary facilities free of cost. Seller shall notify the Buyer through e-mail about readiness of goods for pre-dispatch inspection and Buyer will notify the Seller about the Authorized Representative/ Nominated External Inspection Agency and the date for testing. The goods would be dispatched to consignee only after clearance in pre-dispatch inspection. Consignee's right of rejection as per GTC in respect of the goods finally received at his location shall in no way be limited or waived by reason of the goods having previously been inspected, tested and passed by Buyer/ Consignee or its Nominated External Inspection Agency prior to the goods' shipment. While bidding, the sellers should take into account 7 days for inspection from the date of email offering the goods for inspection. Any delay in inspection beyond 7 days shall be on the part of the buyer and shall be regularised without Liquidated Damages.

When there is requirement of submission the advance sample, the seller shall inform the buyer promptly through emails about the date of submission of sample to the buyer nominated Inspection agency.

4. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

ESIC FUND ACCOUNT NO-02
payable at
RANCHI

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

5. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

6. **Warranty**

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 5 years for the offered products at the rate not more than 10 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Biannually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of

5

(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 5% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

7. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Annexure -I

Specifications for High Definition Laparoscopy Surgery System/4K Laparoscopy Surgery equipment

4Ktelescope/Laparoscope

1. 4K compatible telescope/laparoscope with wide field of vision having 5.0 mm outer diameter with 290 mm or more working length for white light and near infrared application.

Direction of view:30degree1scope

2. 4K compatible telescope/laparoscope having 10 mm outer diameter with 310 mm or more working length for white light and near Infrared applications. Direction of view:0 and 30 degree:1 scope each-Laparoscope/telescope should be fully autoclavable

Optimized glass fibre arrangements to ensure a uniform image brightness of highest quality.

3. Field of view 70-90 degree or more for 5.0 mm & 10 mm laparoscope.-Compatible with all camera heads/couplers.

4. Scope should be of same make as the camera.
5. Should have high resolution and less chromatic aberration
6. Completely distortion free
7. Able to provide 4k white light and ICG images
8. Homogeneous light distribution in the peripheral region
9. Eyepiece type connection for universal compatibility

VIDEOPROCESSOR

1. System should have resolution of 3840x2160 for 2D/3D i.e 3D in 4k resolution(3840x2160) and 2D in 4k resolution(3840x2160) and

d aspect ratio of 16:9 for 3D and 4k video processing (Institution can opt for 2D in 4K resolution or 3D in 4K resolution upgradable system as per their requirement)

2. Should have optical image enhancement for visualization of capillary vessels and fine patterns in the superficial layer of mucosa

3. System should be compatible for both white light and ICG/IR visualization and should have multiple modes of ICG/IR visualization like overlay mode, monochromatic imaging and intensity gain

4. Picture in picture of visualization modes with Standard and Optical Contrast Differentiation.

5. Automatic adjustment of light intensity of light source and controlled from Camera head/console

6. Outputs: All compatible outputs should be there for 4K resolution and DVI for HD resolution.

7. It should have chip on tip technology.

8. It should be focus free.

9. Should have a touch panel control/button press control

10. Should have facility for color customization such to support a range of procedures.

11. Should have digital signal outputs for 4k and HD

12. Should have electronic zoom options up to 2.0x and more in 4k.

LED Light source with fibre optic cable

1. Should be powerful LED light source able to provide white light & ICG light visualization in full color overlay mode

2. Should have lumen > 2000

3. Lamp life of minimum 5 years or 30,000 or more working hours

4. 4-5mm fibre optic cable and 300 cm long

5. Should have facility to select various fluorescence/IR modes such as ICG only ICG, Full color overlay ICG Mode

4K Camera Head

1. Should be CMOS true 4K Camera Head

2. Should have 2 or more programmable head buttons with facility to assign more than 2 functions on single button

3. Should provide 3840x2160 and above, 4k and overlay ICG image.

4. Should be autoclavable /ETO/PLASMA Sterilizer compatible
5. Should have focal length of 18.0 mm or more.
6. Should be compatible with all size of telescopes.
7. Should have automatic shutter & microprocessor controlled automatic gain control
8. Camera head should be able to perform both white light and near infrared applications.
9. Camera Control Function - should be able to change camera setting by console and Camera Head.
10. Camera head cable should be minimum 12 feet long.
11. Camera system should be certified by USFDA / European CE / ICMED 9001 & 13485 (Indian Certification for Medical Devices)/BIS.

4K Medical grade Monitor (55 inch or more)

1. Should have LCD/LED Backlit monitor with capability to show UHD definition resolution of 3840x2160 or more capable of displaying 2D in 4K resolution and 2D in full HD resolution
2. Should have 16:9 aspect ratio
3. Should have multi format display format-picture in picture and picture-in-picture and flip pattern to rotate the image.
4. Should have various 4k & HD input/output terminals including 3G/HD/SD SDI,DVI-D,BNC(x5)/HDMI/DP
5. Monitor should be providing a higher contrast image and less color blurring
6. Must be medical grade monitor
7. Monitor should support picture size(Diagonal) 55 inches or more

Image Management System(Recorder)

1. A high definition / 4k video recorder system with single channel Simultaneous recording facility(medical grade 4k recording with CE/BIS/FDA certificate).
2. Preferably should have a LCD Touch panel Display.
3. Should be able to record a full HD imaging recording
4. Should have internal/external hard disk drive of 1TB or more+2TB external.

5. The recorder should have two inputs of DVI, HD-SDI, Composite video & S-Video for recording.

6. It Should have a feature of real time in-procedure recording capability besides at the end procedure DVD burning.

7. Such a system should also offer a simple video editing capability for easy separation, slicing & deletion of un-necessary parts within a stored video sequence from the on console touch screen itself.

8. Should be capable of offering an easy way to input patient data viz demographic, clinical & diagnostic details and in turn linking the same to the stills & video sequences planned to be archived into a folder format for efficient information management from the touch screen.

9. Should support file format for

a) Images–Bitmap(BMP),JPEG,TaggedImageFileFormat(TIFF)

b) Videos-MPEG4,AVC/H.264

c) Network Integration- System should be capable of integration to the FTP/CIFS/DICOM (C-STORE) Servers

Co2 Insufflator

1. Co2 insufflator should be high flow insufflator which have inbuilt high pressure regulator & heater.

2. It should maintain abdominal pressure during laparoscopic surgery even in case of suction.

3. Warm gas should be supplied by unit to prevent fogging.

4. Maximum flow should be 40 Litre per Minute (LPM) or more.

5. It should be able to show remanent and consumed cylinder status

6. It should have alarm for over pressure, tube obstruction, insufficient supply of gas.

7. Unit should be advance microcontroller microprocessor based

8. It should have display of set pressure, actual pressure, flow rate & volume and delivered pressure.

9. It should be supplied with connecting tube to cylinder & Camera system

10. It should have Power cable-1 (length 3 metre or more)

11. System should be certified with USFDA/European CE /ICMED 9001 & 13485/BIS

General Terms & Conditions

1. System should be supplied with trolley

2. Core equipment like 4K laparoscope, Camera Control Unit, Camera Head, Monitor, light source, 4K scopes, insufflator should be from same make.

3. Cost of consumables not covered under warranty/CMC should also be quoted separately.

rately

4. System should have 05 years Warranty followed by 05 years CMC for all components including camera Head.
5. Any upgradation of software within warranty and CMC period of equipment will be done free of cost by OEM/Vendor.

Annexure -II

Laparoscopy Hand instruments List (compatible with the system):

S.No	Name	Qty
1	VERESS Pneumoperitoneum Needle with spring loaded blunt stylet ,LUE R- lock, length 13 cm &15 cm	02
2	Reusable Trocar, size 11 mm, consisting of Trocar only, with pyramidal tip Cannula without valve, with insufflation stop- cock, length 10.5 cm Multifunctional Valve, three part dismantle. With Washers pack of 10 pcs	04
3	Reusable Trocar, size 11 mm, consisting of: Trocar only, with blunt tip. Cannula without valve, with insufflation stop- cock, length 10.5 cm Multifunctional valve, size 11 mm	02
4	Trocar, size 11 mm, consisting of Trocar only with blunt tip Cannula without valve with insufflation stopcock, with 2 flanges for fixation of sutures length 13 cm Automatic valve Sliding Cone	02
5	Trocar, size 6 mm, consisting of Trocar only, with pyramidal tip Cannula without valve, with insufflation stop- cock, length 10.5 cm Multifunctional Valve, Three part dismantle. With washers pack of 10 pcs	05
6	Trocar 6 mm diameter with thread and rotating insufflations, should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. The working length of the cannula should be 105 mm. Telescope Stopper	02
7	Reducer 11/5mm	02
8	Reduction Sleeve, 11/5mm	02
9	Injection Needle, LUER-lock, diameter 1.2mm, size 5mm, length 36cm	02
10	Three piece laparoscopic KELLY Dissecting and Grasping Forceps, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, double action jaws short, consisting of: Plastic Handle, without ratchet, Outer Tube insulated, Forceps Insert, Three part dismantle	02

11	Three-piece laparoscopic KELLY Dissecting and Grasping Forceps, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, double action jaws long, consisting of: Plastic Handle, without ratchet, with larger contact area. Outer Tube insulated, Forceps Insert, Three part dismantle	02
12	Three-piece laparoscopic Dissecting and Grasping Forceps, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, right angled, double action jaws, consisting of: Plastic Handle, without ratchet, with larger contact area, Outer Tube insulated, Forceps Insert, Three part dismantle	02
13	Three piece laparoscopic Dissecting- and Grasping Forceps, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, alligator jaws, double action jaws, consisting of: Plastic Handle with ratchet , with larger contact area, Outer Tube insulated, Forceps Insert, Three part dismantle	02
14	Three piece laparoscopic REDDICK-OLSEN dissecting and Grasping Forceps, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, heavy, double action jaws, consisting of: Plastic Handle, with ratchet Outer Tube insulated, Forceps Insert, Three part dismantle-03	02
15	Three piece laparoscopic CROCE-OLMI Grasping Forceps, rotating, dismantling, insulated, with connector pin for unipolar coagulation, with LUE R-Lock connector for cleaning, single action jaws, atraumatic, fenestrated, curved, size 5 mm, length 36 cm consisting of: Plastic Handle, with ratchet , Metal Outer Sheath insulated Forceps Insert, Three part dismantle	02
16	Three piece laparoscopic BABCOCK Grasping Forceps, size 5 mm, length 36 cm, with connector pin for unipolar coagulation, atraumatic, jaws with multiple teeth, fenestrated jaws, consisting of: , Plastic Handle, with ratchet, with larger contact area. Outer Tube insulated, Forceps Insert , Three part dismantle	02
17	Three piece laparoscopic Grasping Forceps, 2 x 4 teeth, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, double action jaws, consisting of: • Plastic Handle with ratchet Outer Tube insulated, Forceps Insert, Three part dismantle	02
18	18 Three-piece laparoscopic Bowel Grasper, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, double action jaws, consisting of: Plastic Handle, without ratchet, with larger contact area Outer Tube insulated., Forceps Insert, Three part dismantle	02
19	Three piece laparoscopic Grasping Forceps, rotating, with connector pin for Unipolar coagulation, size 5 mm, length 36 cm, atraumatic, fenestrated,	02
	single action jaws, consisting of: , Plastic Handle, with ratchet , with larger contact area, Outer Tube insulated, Forceps Insert, Three part dismantle	

20	20 BLAKESLEY dissecting and biopsy forceps, 5mm, single action jaws, plastic handle with ratchet and insulated forcep insert	02
21	21 RoBi KELLY Dissecting and grasping forceps, with connector pin for BIPOLAR coagulation, size 5mm length 36 cm, double action jaws, with ring handle, outer sheath, forcep insert	02
22	22 METZENBAUM Scissors, rotating, dismantling, insulated, with connector pin for unipolar coagulation, with LUER-Lock connector for cleaning, double action jaws, curved, length of blades 15 mm, size 5 mm, length 36 cm consisting of: Plastic Handle, insulated, without ratchet Metal Outer Sheath insulated, Scissors Insert, Three part dismantle	02
23	single use scissor insert Pack of 10 (IX10)	02
24	Three piece laparoscopic Hook Scissors, rotating, size 5 mm, length 36 cm, jaws not crossing with connector pin for unipolar coagulation, single action jaws, consisting of: Plastic Handle, without ratchet, with larger contact area Outer Tube insulate, Forceps Insert, Three part dismantle	02
25	Suction and Irrigation Tube ,anti-reflex surface with two-way stopcock, for single hand control, size 5 mm, length 36 cm With trumpet Valve	02
26	Suction or Irrigation Tube, with protection basket ,with trumpet valve, size 10 mm length 36 cm	02
27	Silicone tubing set for use with suction irrigation with trumpet valve	02
28	Coagulating and Dissecting Electrode, spatula-shaped, blunt with connector pin for unipolar coagulation, size 5 mm, working length 36 cm	02
29	Coagulating and Dissecting Electrode, L-shaped, blunt with connector pin for unipolar coagulation, size 5 mm, working length 36 cm	02
30	Three-piece laparoscopic needle holder, ergonomic pistol grip with disengageable ratchet, ratchet release on the left side, straight jaws, with tungsten carbide insert 5 mm, length 33 cm	02
31	Unipolar HFCord with 4mm Bananaplug and length 300cm	02
32	Clip Applicator, for use with Titanium-Clips LT-300 (medium large), dismantling, rotating, with ratchet to lock the jaw part holding the clip, size 10 mm, length 36 cm, consisting of: Metal Handle, with ratchet. Metal Outer Tube Insert, Three parts dismantle	02
33	Clip Applicator ,for use with Titanium-Clips L T-400(-large), dismantling, rotating, with ratchet to lock the jaw part holding the clip, size 10 mm, length	02

	36cm, consisting of: Metal Handle, with ratchet. Metal Outer Tube Insert, Three parts dismantle	
--	---	--

34	Claw Forceps, rotating, size 10 mm, length 36 cm, 2 x 3 teeth, single action jaws, consisting of: Metal Handle, with MANHES style ratchet Outer tube insulated, Forceps Insert, three part dismantle	02
35	Three piece laparoscopic autoclavable forceps for retrieving gallstones , double action 1 jaws with 360 degree rotational sheath, size 10 mm, length 36 cm long, ergonomic metal handle with plastic finger rings with larger contact area at the finger ring to avoid pressure sores, can be dismantled with the press of a button	02
36	BipolarHFCConnectingCablelength300cm,compatiblewithallstandard cauteries	02
37	FascialClosureInstrumentforsubcutaneousligatureoftrocarincisions	02
38	Tray for sterilization of laparoscopicinstruments	02
39	Formalin chamber for keeping laparoscopic instruments	05
40	Knotpusher,5mm	02
41	ReusableEndoscopicpolymerclipapplier,Largesizewithshaftsize33cm and shaft diameter 10 mm	01
42	Polymerclips-Largesize30cartridges	01

- Ø **Experience Certificate** - Bidder should have minimum of 3 years' experience (F.Y . 2022-23, 2023-24 & 2024-25) in supplying same or similar category equipment s/items to any Central / State Govt Organization / Autonomous body/ PSU . Bidder has to submit copies of relevant Contract copy and delivery acceptance certificate like CRAC in support of having supplied some quantity during each of the financial year.
- Ø **Past Performance** - The Bidder should have supplied same or similar Category Products for 50% of bid Value, in at least one of the last three Financial years (FY 2022-23,2023-24 & 2024-25) to any Central / State Govt Organization / Autonomous body/ PSU. Bidder shall submit Copies of relevant Contract copy with CRAC.
- Ø **Turnover Criteria** - The bidder shall submit the copy of Audit P&L Account and Balance sheet (F.Y. 2021-22, 2022-23& 2023-24) or the Turnover certificate issued by the Chartered Accountant/Cost Accountant as per GeM instructions.
- Ø Certificate required for ATC
- Ø **EMD Exemption Certificate** - Under MSE category, only manufacturers for goods are eligible for exemption from EMD. Traders are excluded from the purview of this, hence proper "MSE certificate" must be submitted.
- Ø **EMD** - Scanned Copy / proof of the DD in favor of ESIC Fund Account No-2 payable at Ranchi, wherever applicable.

Ø **Make in India Certificate** – This must be submitted on the letter head of the OEM

Ø **OEM Authorization** – If the bidder is an OEM, then he must submit a self declaration on his letter head in this regard. In case bidder is not OEM, the OEM Authorization certificate issued by the OEM in the name of bidder shall be submitted with other details.

(1) Reverse Auction shall be conducted first 50 % of Technically qualified bidders .

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to

such representations.

Additional Clause For Comprehensive Maintenance Charges

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC. Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1
CMC charges for 2nd year after warranty period- Percentage to be indicated- A2
CMC charges for 3rd year after warranty period - Percentage to be indicated- A3
CMC charges for 4th year after warranty period - Percentage to be indicated- A4
CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C – Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 25% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of

CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है / This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा। In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---